



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
DIVISION OF PUBLIC HEALTH, STATE LABORATORY OF PUBLIC HEALTH

ROY COOPER  
GOVERNOR

MANDY COHEN, MD, MPH  
SECRETARY

DANIEL STALEY  
DIRECTOR

SCOTT J. ZIMMERMAN, DRPH, MPH, HCLD (ABB)  
LABORATORY DIRECTOR

# Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Heather Colvin, B.S  
Laboratory Improvement Consultant

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** April 17, 2018

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**GC-Lect:** Lot Number: **8068568** Expiration Date: **06/16/2018**  
Along with **Bio-Bag™ Type C:** Lot Number: **7128668** Expiration Date: **02/28/2019**  
and with **Bio-Bag™ Type C:** Lot Number: **7296619** Expiration Date: **06/30/2019**

**Summary of NCSLPH QC results for this Lot #:**

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
<i>N. gonorrhoeae</i> ATCC# 43070 - Growth	<i>P. mirabilis</i> ATCC# 43071 – No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None

BD and Fisher Scientific contacted: April 13, 2018.

\*If your contact information changes, please provide that information to Heather Colvin in Laboratory Improvement at (919) 733-7186 or via e-mail at [heather.colvin@dhhs.nc.gov](mailto:heather.colvin@dhhs.nc.gov). Thank you.